

Référence de votre commande : 4535212352 du 23/02/2026

54 avenue de la Plaine

Z.I. de Rousset

13790 ROUSSET

SAS au capital de 1 000 €

RCS :
SIRET : 82199551100010
N°id CEE : FR68821995511
Téléphone : 04 42 29 06 43
Télécopie : 04 42 29 06 75
Email :
Site :

Facturation

DIVERSEY EUROPE OPERATIO
DE CORRIDOR 4

3621 ZB BREUKELEN

Pays bas

Vid CEE : NL851079623B01

Livraison

H. Essers en Zonen Int.
Transport NV
Henry Fordlaan 8 14 D-E
3600 GENK
Belgique

Nous avons bien reçu votre commande et vous en remercions. Le délai de livraison indiqué ci-dessus n'est donné qu'à titre indicatif. La commande sera exécutée conformément aux "CONDITIONS GENERALES DE VENTES" de SIMA PHARMA. Réserve de propriété : Les marchandises commandées demeurent notre propriété jusqu'au complet paiement de leurs prix en principal frais et accessoires, notamment dans les conditions des articles 121 et 122 de la loi 85-98 du 25 janvier 1985 ; toute condition contraire à la présente clause de réserve de propriété, résultant des conditions générales, des bons de commande, ou plus généralement de tout document quelconque de l'acquéreur, nous est inopposable sans qu'il soit besoin de protestation de notre part. Toute réclamation concernant une livraison doit être formulée au plus tard 8 jours après réception de nos marchandises. A la réception de la marchandise, vous en serez le dépositaire et vous ne pourrez les transformer ou les revendre sans notre accord et avant paiement intégral du prix. Pendant la durée de réserve de propriété, les risques en cas de perte ou de destruction des biens livrés sont à votre charge.

Référence	Désignation	Qté	Reliquat
PRESTAPHARMA 101109135	Commande n°BC001609 du 23/02/2026		
	Prestation de conditionnement pharmaceutique	1,00	
	DI ClearKlens Plus VH5 S 100X0.03I W4449 Lot SPH06052026	80,00	
CSMJTL	Commande Soldée Enlèvement par MJTL, 80 colis sur 5 palettes Poids 581 Kg Certificat de conformité ci-joint	1,00	
CONGES	1 colis livraison chez MEDICAL LAB par Ciblex Poids 0Kg500 FERMETURE POUR LES CONGES Nous vous informons que SIMAPHARMA sera fermé pour les vacances d'été du Samedi 01 Août 2026 au matin jusqu'au Dimanche 23 Août 2026 inclus. Merci d'anticiper vos commandes si nécessaire.	1,00	

SIMA PHARMA	Form	SP-PR-LIB-F-023	Creation date : 12/10/2023
	Certificate of Analysis DIVERSEY for ClearKlens Plus VH5 100x0.03L		Modification date : 21/10/2025
			Index : 04

 ClearKlens Plus VH5 100x0.03L SKU 101109135	
Batch Number	SPH06052026
Production date	05/2026
Expiration Date	04/2028

This is to certify that the above batch of product has been tested and conforms to the manufacturing Quality Control specification for the product.

ClearKlens Plus VH5				
Test	Test Method	Specification	Unit	Results
Appearance	SP-PR-LIB-I-011	Clear colourless to yellow liquid	N/A	Clear colourless to yellow liquid
pH neat (20°C)	PS-CON-I-021	5.0 – 6.0	-	5.3
pH (1% solution in water, 20°C)	PS-CON-I-021	5.4 – 6.4	-	6.0
Specific gravity (20°C)	SP-PR-LIB-I-010	1.010 - 1.050	-	1.037
Anionic substance content (%)	SP-PR-LIB-I-018	15.50 - 16.50	-	16.38

Control of the Sterility of the product :


Test	Test Method	Specification	Conformity	Reference number
Sterilization	Gamma Irradiation	25.0 – 40.0 (kGy)	C	2023-31113A
Sterility certificate	Membrane filtration sterility testing EP 11.0 2.6.1	No Growth	C	26/CBH/STE/093


The analysis results above could change over time in function of the storage temperature. It is imperative to store the products according to the recommended conditions indicated in the Safety Data Sheet.

SIMA PHARMA	Form	SP-PR-LIB-F-023	Creation date : 12/10/2023
	Certificate of Analysis DIVERSEY for ClearKlens Plus VH5 100x0.03L		Modification date : 21/10/2025
			Index : 04

 ClearKlens Plus VH5 100x0.03L SKU 101109135	
Batch Number	SPH06052026
Production date	05/2026
Expiration Date	04/2028

We certify that, except for the exceptions or deviations listed above, the quoted supply has been manufactured and tested in accordance with the requirements of the specifications in force

SIMA-PHARMA Control (Responsible for the production process of the final product) :	Stamp and signature
Name: E. OACQUIN Position: General Manager Produced on behalf of Diversey at : Sima Pharma: Z.I. de Rousset / Peynier, 54 Av. de la Plaine, 13790 Rousset, France Date : 26/05/26	 SIMA PHARMA 54 Avenue de la Plaine, ZI 13106 ROUSSET Cedex Tél. 04 42 29 06 43 - Fax 04 42 29 06 75 SAS au capital de 1 000 € RCS AIX 821 995 511

DIVERSEY Control	Stamp and signature
Name : Position: Made at : Diversey Europe Operations B.V., A Solenis company, De Corridor 43621 ZB Breukelen, Netherlands On :	Céline Lovato Global Life Sciences Director Signed by:  7459206AC8ED40A... 30th June 2026

End of certificate of analysis

For further information, please contact Diversey at lifesciences@solenis.com

STERIS: Certificate Of Processing

Prepared For **SIMA PHARMA (11582)**
Gamma Process Run ID **2023-31113A**

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
320340P - CLEARLENSE PLUS VH5 30 ML	SPH06052026	4	Each

Validation Reference Number: 2023-12537A

Processing Run Start Date 29-May-2026 2:24 AM

Processing Run End Date 29-May-2026 8:45 AM

Specified Dose Range (kGy)	25.0 - 40.0	Calculated Min Dose (kGy)	26.7
Reference Dose Range (kGy)	29.3 - 32.7	Calculated Max Dose (kGy)	39.7

PO Number: FBC03311 - 27/05/2026

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Gamma Process Run Approval authorized by STERIS

DateTime Esigned 29-May-2026 2:31 PM

Operating facilities are in compliance with applicable local/state and federal regulations providing services under a certified quality system which meets the following requirements (if applicable): FDA QMSR, ISO 9001, ISO 13485 current certified version, national pharmaceutical GMP (EU GMP, Commission Directive 2017/1572, Veterinary Regulation (EU) 2019/6, etc.) and is in alignment with ISO 11137 current certified version. STERIS AST certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used. The product has been processed as per the requirements of the Customer Agreement and/or Customer Specification.

Processing Location

Synergy Health Marseille SAS,
a STERIS Company
MIN 712 Les Arnavaux
13323 Marseille Cedex 14
France
Phone: +33 (0) 491-214-214

SIMA PHARMA
54 Avenue de la Plaine, ZI
13106 ROUSSET Cedex
Tél. 04 42 29 06 43 - Fax 04 42 29 06 75
SAS au capital de 1 000 €
RCS AIX 821 995 511

Contrôle ok

NOUVEAU i

29/05/2026



Document ID: 13599